

VERIFICATION OF CALIBRATION REPORT*of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	<u>130180</u> ✓
Name	<u>A.W. Anderson</u>	ID#	<u>2467</u> ✓
Date	<u>11/4/10</u> ✓		
A Agency	<u>Seldovia PD</u>	Phone #	<u>234-7640</u>
Instrument Location	<u>Seldovia PD</u>		
B Alco S/N	<u>X301530</u> ✓	Target Value	<u>.078</u> ✓
High Pressure	<u>500 PSI</u>		
Alco Test Values	<u>.072</u> ✓	<u>.073</u> ✓	
1st Alco		2nd Alco	
Signature	<u>AW Anderson</u>		

(OVER)

SMB
12/9/10

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 23RD day of DEC, 2010.

Daniel L. Nicholas
 Daniel L. Nicholas
 Notary Public, State of Alaska
 Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130180

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

NOVEMBER 04, 2010

OPERATOR'S NAME:

ANDERSON AM

OPERATOR'S NUMBER: 2467

SUBJECT'S LAST NAME:

SIXTY DAY VERIFICATION

SUBJECT'S FIRST NAME/MI:

N/A

O.L. #: 1234567890

DEPT/AGENCY: SOU1

CASE/REPORT: N/A

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301530

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 28.97 in
ALCO TARGET .075 09:46
BLANK TEST .000 09:47
INTERNAL STANDARD VERIFIED 09:47
ALCO TV 28.97 in .072 09:48
BLANK TEST .000 09:49
SUBJECT SAMPLE .000 09:49
BLANK TEST .000 09:50
ALCO TV 28.97 in .073 09:50
BLANK TEST .000 09:51

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180

NOVEMBER 04, 2010
TIME 09:51

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 42c
BREATH TUBE: 42c
BAROMETER: 28.97 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#\$%^&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~